



Ohio Department of Health
Alcohol and Drug Testing
Subject Test Report

Subject Information

Revised 10-2011

TEST DATE 10/27/2013	NAME WILLIAMS, AARON	DRIVERS LICENSE NUMBER [REDACTED]	STATE OH
CURRENT STREET ADDRESS (As Verified by Officer) 3021 COOPER RD		SOCIAL SECURITY #	
CITY CINCINNATI	STATE OH	ZIP CODE 45241	DATE OF BIRTH 03/23/1994
		AGE 19	SEX M



Arrest Information

ARRESTING OFFICER WAGERS, MATTHEW	AGENCY OXFORD POLICE DEPARTMENT
TESTING OFFICER FENING, LARA	AGENCY OXFORD POLICE DEPARTMENT
	ODH CERTIFICATION# 13871
TIME OF VIOLATION 02:13	TIME FIRST OBSERVED 02:57
	IMPLIED CONSENT WARNING Y
	TIME OF TEST 03:22

Test Information

INSTRUMENT SERIAL # 80-004494	TEST SITE # OXFORD PD	DATE OF LAST CERTIFICATION 10/22/2013	CERTIFICATION SOLUTION # ODH-0020
CERTIFICATION BOTTLE # 0517	TARGET VALUE 0.099 g/210L	CERTIFICATION AVERAGE 0.1010	CERTIFICATION STANDARD DEVIATION 0.0007
Test	BrAC (g/210L)	Time	DRY GAS STANDARD
Air Blank	0.000	03:26	LOT # 07113100A2
Diagnostic	VAC/OK	03:26	TANK # 031
Air Blank	0.000	03:27	TARGET: 0.100 g/210L
Dry Gas Control	0.099	03:27	EXPIRATION DATE: 09/01/2015
Air Blank	0.000	03:28	
Subject Sample 1	0.140	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:32	
Subject Sample 2	0.122	03:32	
Air Blank	0.000	03:33	
Dry Gas Control	0.100	03:33	
Air Blank	0.000	03:34	

BrAC = 0.122 g/210L

	10/27/2013
ARRESTING OFFICER'S SIGNATURE	DATE
	10/27/2013
TESTING OFFICER'S SIGNATURE	DATE